DOUBLE MEDICAL
Dedicated to Rehabilitation

Non-Sterile Surgical Mask
With ear loop 17.5cmx9.5cm



CE 50PCS EN 14683:2019 Type IIR

Product composition

Non-sterile Surgical Mask is composed of mask body,nose clip and mask band. The mask body has three layers, with inner and outer layer made of non-woven fabric and middle layer made of meltblown. The product is provided non-sterile.

Instruction for use:

1. Open the package, take out the mask and hold both sides of the mask band, with the nose clip facing out.
2. Cover the mouth, nose and chin of the mask with the mask band behind the ear.
3. Adjust the nose clip to fit closely to the bridge of the nose.
4. Adjust the position of the mask to make it comfortable to wear.
5. Wear the mask correctly according to the front and back signs. Hands off from the inside of the mask.

DOUBLE MEDICAL
Dedicated to Rehabilitation

Non-Sterile Surgical Mask
With ear loop 17.5cmx9.5cm

DOUBLE MEDICAL
Dedicated to Rehabilitation

Non-Sterile Surgical Mask
With ear loop 17.5cmx9.5cm



CE 50PCS EN 14683:2019 Type IIR

Information supplied by manufacturer

【Lot number】 See package code or label.
【Manufacturing date】 See package code or label.
【Effective period】 18 months
【Manufacturer】 Double Medical Technology Inc.
【Address】 No. 18, Shanbianhong East Road, Haicang District, 361026 Xiamen, Fujian, PEOPLE'S REPUBLIC OF CHINA
【Tel】 +86 592 6087101
【After-sale service】 Double Medical Technology Inc.
EC REP Lotus NL B.V.
Koningin Julianaplein 10, le Verd, 2595AA, The Hague, Netherlands
【Contract Manufacturer】 Xiamen Double Micro Surgical Precision Technology Co.,Ltd.
【Address】 Room 2219 & 2220, Workshop Building, No.18, Shanbianhong East Road, Haicang District, Xiamen, P.R. China

Made in China
Product registration code
闽药监械生产许20200577号（临时）
Production license code
闽械注准20202140261（临时）

6 973276 520012

【Lot number】
【Manufacturing date】

Applicable scope

For clinical personnel to wear during invasive operation, covering the user's mouth, nose and jaw, providing a physical barrier to prevent the direct passage of pathogens, microorganisms, body fluids, particles, etc.

Contraindications

Not found yet.

Storage and Transportation

1. Transportation: the product shall be transported by general means, which shall prevent violent impact, collision and rain splashing.
2. Storage: the packaged product shall be stored in a room with relative humidity not exceeding 80%, free from corrosive gases and good ventilation, and away from pollution sources.

Warnings and Precautions:

1.Read the instructions for use carefully and avoid hand contact with inner side of the mask.
2.Non-sterile masks are not used in areas with strict microbiological control.
3.Check the package before use, don't use if it's broken.
4.The mask is disposable and shall not be reused.
5.In case of liquid spillage, broken mask, dampness and significant increase in respiratory resistance, replace the mask in time;
6.After-use mask shall be disposed according to the requirements of the hospital or the environmental protection department.

Graphicinterpretation:

	Do not re-use		Caution
	Do not use if package is damaged		Non-sterile
	Keep away from sunlight		Keep dry